



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

TRADEPORT ELECTRONICS CALIBRATION LABORATORY  
1750 Steeles Ave. West, Unit 5  
Concord, Ontario L4K 2L7 CANADA  
Barry Conway Phone: 905 660 3797

### CALIBRATION

Valid To: July 31, 2024

Certificate Number: 2450.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,6</sup>:

#### I. Electrical – DC / Low Frequency

Parameter/Equipment	Range	CMC <sup>2,5</sup> (±)	Comments
DC Voltage <sup>3</sup> – Generate	(0 to 330) mV 330 mV to 3.3 V (3.3 to 33) V (33 to 330) V (330 to 1000) V  (0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V  (0 to 0.202) V (0.202 to 2.02) V (2.02 to 20.2) V (20.2 to 220) V (200 to 1025) V  (> 1 to 10) kV (10 to 20) kV (20 to 30) kV (30 to 40) kV	16 µV/V + 1.6 µV 8.5 µV/V + 9.6 µV 9.3 µV/V + 180 µV 14 µV/V + 1 mV 14 µV/V + 3.4 mV  8.0 µV/V + 0.68 µV 7.0 µV/V + 1.4 µV 7.0 µV/V + 6.5 µV 7.0 µV/V + 12 µV 8.0 µV/V + 140 µV 9.0 µV/V + 1.5 mV  15 µV/V + 2.7 µV 9.0 µV/V + 2.6 µV 8.0 µV/V + 0.16 mV 12 µV/V + 0.31 mV 12 µV/V + 9.0 mV  0.58 % + 25 V 0.58 % + 26 V 0.58 % + 27 V 0.58 % + 30 V	Fluke 5520A  Fluke 5700A  Transmille 4010  Pintek HVC-801

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
DC Voltage <sup>3</sup> – Measure	(0 to 100) mV (0 to 1) V (0 to 10) V (0 to 100) V (0 to 1000) V  100 V to 15 kV  (0.5 to 6) kV (> 6 to 10) kV  (> 10 to 40) kV (40 to 70) kV (70 to 100) kV	11 $\mu$ V/V + 1.0 $\mu$ V 10 $\mu$ V/V + 1.0 $\mu$ V 10 $\mu$ V/V + 2.1 $\mu$ V 12 $\mu$ V/V + 34 $\mu$ V 24 $\mu$ V/V + 0.13 mV  3.9 %  0.035 % + 0.58 V 0.035 % + 5.8 V  0.058 % + 5.8 V 0.058 % + 58 V 0.058 % + 0.12 kV	Agilent 3458A  Tektronix P6015A with 1 M $\Omega$ impedance oscilloscope  Vitrek 4700  Vitrek 4700 & Vitrek HVL-100 probe

Parameter/Equipment	Range	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
DC Current <sup>3</sup> – Generate (cont)	(0 to 202) $\mu$ A (0.2 to 2.02) mA (2 to 20.2) mA (20 to 202) mA  (0.2 to 2.02) A (2 to 20.2) A (20.2 to 30) A  (> 20 to 50) A (50 to 100) A (100 to 500) A (500 to 1000) A	0.01 % + 12 nA 0.050 % + 79 nA 0.050 % + 0.37 $\mu$ A 0.05% + 3.7 $\mu$ A  0.013 % + 60 $\mu$ A 0.030 % + 1.4 mA 0.050 % + 1.5 mA  0.25 % + 0.20 A 0.25 % + 0.40 A 0.25 % + 1.0 A 0.25 % + 1.8 A	Transmille 4010  Transmille 4010 & Fluke 5500A/COIL
DC Current <sup>3</sup> – Measure	(0 to 100) $\mu$ A 100 $\mu$ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A  (1 to 3) A (3 to 10) A  (10 to 20) A	25 $\mu$ A/A + 2.1 nA 25 $\mu$ A/A + 17 nA 25 $\mu$ A/A + 0.17 $\mu$ A 40 $\mu$ A/A + 2.4 $\mu$ A 120 $\mu$ A/A + 50 $\mu$ A  0.016 % + 4.4 mA 0.039 % + 4.4 mA  0.21 % + 9.7 mA	Agilent 3458A  Fluke 8846A  Agilent 3458A & 50 A/50 mV current shunt
Resistance <sup>3</sup> – Generate	(0 to 11) $\Omega$ (11 to 33) $\Omega$ (33 to 110) $\Omega$ (110 to 330) $\Omega$ 330 $\Omega$ to 1.1 k $\Omega$ (1.1 to 3.3) k $\Omega$ (3.3 to 11) k $\Omega$ (11 to 33) k $\Omega$ (33 to 110) k $\Omega$ (110 to 330) k $\Omega$ 330 k $\Omega$ to 1.1 M $\Omega$ (1.1 to 3.3) M $\Omega$ (3.3 to 11) M $\Omega$ (11 to 33) M $\Omega$ (33 to 110) M $\Omega$ (110 to 330) M $\Omega$ 330 M $\Omega$ to 1.1 G $\Omega$	31 $\mu$ $\Omega$ / $\Omega$ + 0.79 m $\Omega$ 23 $\mu$ $\Omega$ / $\Omega$ + 1.2 m $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 1.5 m $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 0.024 $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 0.0059 $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 0.030 $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 0.083 $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 0.30 $\Omega$ 22 $\mu$ $\Omega$ / $\Omega$ + 0.90 $\Omega$ 25 $\mu$ $\Omega$ / $\Omega$ + 2.7 $\Omega$ 25 $\mu$ $\Omega$ / $\Omega$ + 8.4 $\Omega$ 47 $\mu$ $\Omega$ / $\Omega$ + 48 $\Omega$ 0.010 % + 200 $\Omega$ 0.019 % + 2.7 k $\Omega$ 0.039 % + 9.4 k $\Omega$ 0.23 % + 110 k $\Omega$ 1.2 % + 1.3 M $\Omega$	Fluke 5520A

Parameter/Equipment	Range	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
Resistance <sup>3</sup> – Generate (cont)	(0 to 100) $\Omega$ (100 to 330) $\Omega$ 330 $\Omega$ to 1.1 k $\Omega$ (1.1 to 3.3) k $\Omega$ (3.3 to 10) k $\Omega$ (10 to 33) k $\Omega$ (33 to 100) k $\Omega$ (100 to 330) k $\Omega$ 330 k $\Omega$ to 1 M $\Omega$ (1 to 3.3) M $\Omega$ (3.3 to 10) M $\Omega$ (10 to 33) M $\Omega$ (33 to 100) M $\Omega$ (100 to 330) M $\Omega$ 330 M $\Omega$ to 1.0 G $\Omega$	0.010 % + 50 m $\Omega$ 0.010 % + 50 m $\Omega$ 0.010 % + 51 m $\Omega$ 0.010 % + 94 m $\Omega$ 0.010 % + 0.14 $\Omega$ 0.010 % + 0.82 $\Omega$ 0.010 % + 1.4 $\Omega$ 0.010 % + 43 $\Omega$ 0.010 % + 53 $\Omega$ 0.010 % + 0.17 k $\Omega$ 0.010 % + 0.30 k $\Omega$ 0.010 % + 9.3 k $\Omega$ 0.050 % + 0.10 M $\Omega$ 1.0 % + 0.21 M $\Omega$ 2.0 % + 0.70 M $\Omega$	Transmille 4010
Resistance – Generate, (2-wire and 4-wire) Fixed Points	0 $\Omega$ 1 $\Omega$ 1.9 $\Omega$ 10 $\Omega$ 19 $\Omega$ 100 $\Omega$ 190 $\Omega$ 1 k $\Omega$ 1.9 k $\Omega$ 10 k $\Omega$ 19 k $\Omega$ 100 k $\Omega$ 190 k $\Omega$ 1 M $\Omega$ 1.9 M $\Omega$ 10 M $\Omega$ 19 M $\Omega$ 100 M $\Omega$  0.001 $\Omega$ 0.01 $\Omega$ 0.1 $\Omega$ 1 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$	56 $\mu\Omega$ 120 $\mu\Omega$ 220 $\mu\Omega$ 370 $\mu\Omega$ 720 $\mu\Omega$ 2.0 m $\Omega$ 3.0 m $\Omega$ 16 m $\Omega$ 32 m $\Omega$ 120 m $\Omega$ 290 m $\Omega$ 1.7 $\Omega$ 2.7 $\Omega$ 25 $\Omega$ 51 $\Omega$ 490 $\Omega$ 1.7 k $\Omega$ 16 k $\Omega$  0.15 $\mu\Omega$ 0.94 $\mu\Omega$ 16 $\mu\Omega$ 0.16 m $\Omega$ 0.73 m $\Omega$ 14 m $\Omega$ 87 m $\Omega$ 1.0 $\Omega$ 23 $\Omega$	Fluke 5700A  KAENAHOB CCCP fixed value resistor

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Resistance <sup>3</sup> – Measure (2-Wire and 4-Wire)	(0 to 10) Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ	18 μΩ/Ω + 0.11 mΩ 15 μΩ/Ω + 1.0 mΩ 13 μΩ/Ω + 7.5 mΩ 13 μΩ/Ω + 75 mΩ 13 μΩ/Ω + 0.75 Ω 18 μΩ/Ω + 8.3 Ω 53 μΩ/Ω + 0.20 kΩ 0.050 % + 0.24 kΩ 0.50 % + 0.24 MΩ	Agilent 3458A

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
AC Voltage <sup>3</sup> – Generate			
(1 to 33) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.062 % + 5.0 μV 0.012 % + 5.0 μV 0.016 % + 5.6 μV 0.078 % + 7.7 μV 0.27 % + 12 μV 0.62 % + 45 μV	Fluke 5520A
(33 to 330) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.023 % + 25 μV 0.011 % + 11 μV 0.012 % + 11 μV 0.027 % + 33 μV 0.062 % + 35 μV 0.16 % + 93 μV	
330 mV to 3.3 V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.023 % + 150 μV 0.012 % + 100 μV 0.015 % + 100 μV 0.023 % + 120 μV 0.054 % + 220 μV 0.19 % + 1.0 mV	
(3.3 to 33) V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.023 % + 3.4 mV 0.012 % + 1.0 mV 0.019 % + 1.1 mV 0.027 % + 1.5 mV 0.070 % + 2.6 mV	

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
AC Voltage <sup>3</sup> – Generate (cont)			
(33 to 330) V	45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.015 % + 12 mV 0.016 % + 13 mV 0.019 % + 14 mV 0.023 % + 31 mV 0.16 % + 49 mV	Fluke 5520A
(330 to 1020) V	45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.023 % + 54 mV 0.019 % + 32 mV 0.023 % + 35 mV	
(1 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.055 % + 4.5 µV 0.021 % + 4.5 µV 0.011 % + 4.5 µV 0.037 % + 4.5 µV 0.085 % + 7.0 µV 0.11 % + 13 µV 0.17 % + 25 µV 0.34 % + 25 µV	Fluke 5700A
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.055 % + 5.1 µV 0.021 % + 5.2 µV 0.011 % + 4.2 µV 0.037 % + 5.1 µV 0.085 % + 7.0 µV 0.11 % + 12 µV 0.17 % + 25 µV 0.34 % + 25 µV	
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.055 % + 13 µV 0.021 % + 8.0 µV 0.011 % + 8.0 µV 0.037 % + 8.0 µV 0.085 % + 25 µV 0.11 % + 25 µV 0.17 % + 35 µV 0.34 % + 80 µV	
(0.22 to 2.2) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.050 % + 130 µV 0.016 % + 55 µV 0.0080 % + 41 µV 0.012 % + 83 µV 0.025 % + 140 µV 0.043 % + 280 µV 0.11 % + 530 µV 0.22 % + 1.6 mV	

Parameter/Range	Frequency	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
AC Voltage <sup>3</sup> – Generate (cont)			
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.050 % + 1.3 mV 0.016 % + 560 $\mu$ V 0.008 % + 440 $\mu$ V 0.012 % + 860 $\mu$ V 0.025 % + 1.3 mV 0.050 % + 3.3 mV 0.13 % + 7.5 mV 0.27 % + 20 mV	Fluke 5700A
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	0.050 % + 13 mV 0.016 % + 5.8 mV 0.0080 % + 5.1 mV 0.022 % + 12 mV 0.050 % + 17 mV	
(220 to 1100) V	(15 to 50) Hz 50 Hz to 1 kHz	0.040 % + 95 mV 0.0080 % + 94 mV	
(0 to 202) mV	(10 to 45) Hz 45 Hz to 1 kHz (1 to 20) kHz (20 to 100) kHz (100 to 500) kHz	0.080 % + 31 $\mu$ V 0.016 % + 31 $\mu$ V 0.020 % + 44 $\mu$ V 0.10 % + 0.15 mV 0.40 % + 0.17 mV	Transmille 4010
(0.2 to 2.02) V	(10 to 45) Hz 45 Hz to 1 kHz (1 to 20) kHz (20 to 100) kHz 100 kHz to 1 MHz	0.050 % + 0.28 mV 0.016 % + 0.25 mV 0.021 % + 0.45 mV 0.065 % + 1.3 mV 0.30 % + 3.3 mV	
(2 to 20.2) V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 100) kHz	0.050 % + 2.6 mV 0.016 % + 2.2 mV 0.021 % + 2.2 mV 0.060 % + 7.2 mV	
(20 to 202) V	(30 Hz to 45) Hz 45 Hz to 1 kHz (1 to 10) kHz (10 to 40) kHz (40 to 100) kHz	0.050 % + 28 mV 0.015 % + 23 mV 0.020 % + 80 mV 0.030 % + 0.14 V 0.20 % + 0.15 V	
(200 to 1020) V	30 Hz to 45 Hz 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz	0.055 % + 0.23 V 0.020 % + 0.13 V 0.025 % + 0.17 V 0.030 % + 0.25 V	

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
AC Voltage <sup>3</sup> – Measure			
(1 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.30 µV/V + 4.1 µV 0.20 µV/V + 3.0 µV 0.30 µV/V + 3.1 µV 1.0 µV/V + 3.1 µV 5.0 µV/V + 9.1 µV 40 µV/V + 66 µV 40 µV/V + 66 µV	Agilent 3458A
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 6) MHz (6 to 8) MHz (8 to 10) MHz	72 nV/V + 7.0 µV 72 nV/V + 6.1 µV 0.14 µV/V + 8.1 µV 0.30 µV/V + 8.1 µV 0.80 µV/V + 37 µV 3.0 µV/V + 70 µV 10 µV/V + 70 µV 15 µV/V + 91 µV 40 µV/V + 0.72 mV 40 µV/V + 0.77 mV 40 µV/V + 0.77 mV 0.015 % + 2.3 mV	
100 mV to 1V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 6) MHz (6 to 8) MHz (8 to 10) MHz	72 nV/V + 64 µV 72 nV/V + 54 µV 0.014 % + 72 µV 0.030 % + 0.13 mV 0.080 % + 0.21 mV 0.30 % + 0.61 mV 1.0 % + 1.7 mV 1.5 % + 1.8 mV 4.0 % + 7.2 mV 4.0 % + 7.7 mV 4.0 % + 7.7 mV 15 % + 25 mV	
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	72 µV/V + 0.71 mV 72 µV/V + 0.54 mV 0.014 % + 0.71 mV 0.030 % + 1.3 mV 0.080 % + 1.6 mV 0.30 % + 5.2 mV 1.0 % + 16 mV	
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.020 % + 9.3 mV 0.020 % + 8.6 mV 0.020 % + 12 mV 0.035 % + 14 mV 0.12 % + 34 mV	

Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
AC Voltage <sup>3</sup> – Measure (cont)			
(10 to 100) V	(100 to 300) kHz 300 kHz to 1 MHz	0.40 % + 35 mV 1.5 % + 35 mV	Agilent 3458A
(100 to 700) V	(1 to 50) Hz 50 Hz to 1 kHz (1 to 10) kHz	0.040 % + 87 mV 0.040 % + 80 mV 0.060 % + 80 mV	
100 V to 15 kV	60 Hz	3.9 %	Tektronix P6015A with $1\text{M}\Omega$ impedance oscilloscope
(0.5 to 5) kV (> 5 to 10) kV	60 Hz	0.14 % + 0.60 V 0.14 % + 29 V	Vitrek 4700
(> 10 to 20) kV (20 to 50) kV (50 to 70) kV	60 Hz	0.14 % + 29 V 0.14 % + 58 V 0.14 % + 0.12 kV	Vitrek 4700 + Vitrek HVL-100 Probe
100 V to 1 kV	400 Hz	0.46 % + 0.13 V	Vitrek 4700
AC Current <sup>3</sup> – Generate			
(29 to 330) $\mu$ A	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.16 % + 0.11 $\mu$ A 0.12 % + 0.10 $\mu$ A 0.097 % + 0.11 $\mu$ A 0.23 % + 0.16 $\mu$ A 0.62 % + 0.22 $\mu$ A 1.2 % + 0.41 $\mu$ A	Fluke 5520A
(0.33 to 3.3) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.16 % + 0.49 $\mu$ A 0.097 % + 0.42 $\mu$ A 0.078 % + 0.41 $\mu$ A 0.16 % + 0.51 $\mu$ A 0.39 % + 0.93 $\mu$ A 0.78 % + 2.8 $\mu$ A	

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
AC Current <sup>3</sup> – Generate (cont) (3.3 to 33) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.14 % + 4.5 µA 0.070 % + 4.3 µA 0.031 % + 4.1 µA 0.062 % + 5.0 µA 0.16 % + 9.0 µA 0.31 % + 21 µA	Fluke 5520A
(0 to 220) µA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.070 % + 28 nA 0.035 % + 24 nA 0.014 % + 21 nA 0.060 % + 42 nA 0.16 % + 82 nA	Fluke 5700A
220 µA to 2.2 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.070 % + 140 nA 0.035 % + 170 nA 0.014 % + 140 nA 0.060 % + 420 nA 0.16 % + 810 nA	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.070 % + 1.5 µA 0.035 % + 1.4 µA 0.014 % + 1.4 µA 0.060 % + 4.3 µA 0.16 % + 8.1 µA	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.070 % + 1.4 µA 0.035 % + 1.4 µA 0.014 % + 1.4 µA 0.060 % + 42 µA 0.16 % + 81 µA	
220 mA to 2.2 A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.065 % + 140 µA 0.075 % + 160 µA 0.85 % + 210 µA	
(33 to 330) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.14 % + 47 µA 0.070 % + 45 µA 0.031 % + 42 µA 0.078 % + 62 µA 0.16 % + 120 µA 0.31 % + 310 µA	Fluke 5520A
(0.33 to 1.1) A	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.14 % + 0.14 mA 0.039 % + 0.13 mA 0.47 % + 0.80 mA 2.0 % + 3.9 mA	

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
AC Current <sup>3</sup> – Generate (cont)			
(1.1 to 3) A	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.14 % + 1.1 mA 0.046 % + 1.0 mA 0.47 % + 1.5 mA 2.0 % + 4.4 mA	Fluke 5520A
(3 to 11) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.047 % + 3.1 mA 0.078 % + 3.1 mA 0.23 % + 4.0 mA	
(11 to 20.5) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.093 % + 6.3 mA 0.12 % + 6.1 mA 0.23 % + 7.5 mA	
(> 20 to 50) A	(50 to 100) Hz (100 to 400) Hz	0.23 % + 0.12 A 0.62 % + 0.12 A	Fluke 5520A & Fluke 5500A/coil
(50 to 100) A	(50 to 100) Hz (100 to 400) Hz	0.23 % + 0.21 A 0.62 % + 0.21 A	
(100 to 500) A	(50 to 100) Hz (100 to 400) Hz	0.23 % + 1.1 A 0.62 % + 1.1 A	
(500 to 1000) A	(50 to 100) Hz (100 to 400) Hz	0.24 % + 2.1 A 0.63 % + 1.9 A	
(20 to 202) µA	(10 to 45) Hz 45 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz	0.20 % + 0.30 µA 0.070 % + 0.23 µA 0.80 % + 0.77 µA 1.6 % + 0.83 µA	Transmille 4010
(0.2 to 2.02) mA	(10 to 45) Hz 45 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz	0.20 % + 0.63 µA 0.060 % + 0.62 µA 0.50 % + 3.3 µA 1.0 % + 6.6 µA	
(2 to 20.2) mA	(10 to 45) Hz 45 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz	0.20 % + 4.0 µA 0.040 % + 3.4 µA 0.25 % + 4.2 µA 0.50 % + 5.0 µA	
(20 to 202) mA	(10 to 45) Hz 45 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz	0.20 % + 58 µA 0.040 % + 54 µA 0.50 % + 70 µA 0.70 % + 0.22 mA	

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
AC Current <sup>3</sup> – Generate (cont)			
(0.2 to 2.02) A	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 30) kHz	0.20 % + 0.41 mA 0.060 % + 0.49 mA 0.50 % + 0.60 mA 0.60 % + 1.1 mA	Transmille 4010
(2 to 30.0) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.080 % + 5.6 mA 0.30 % + 6.6 mA 0.60 % + 6.6 mA 3.0 % + 6.8 mA	
(> 20 to 50) A	(50 to 100) Hz (100 to 400) Hz	0.29 % + 0.11 A 0.79 % + 0.11 A	Transmille 4010 & Fluke 5500A/coil
(50 to 100) A	(50 to 100) Hz (100 to 400) Hz	0.29 % + 0.20 A 0.79 % + 0.20 A	
(100 to 500) A	(50 to 100) Hz (100 to 400) Hz	0.30 % + 1.0 A 0.80 % + 1.0 A	
(500 to 1000) A	(50 to 100) Hz (100 to 400) Hz	0.41 % + 2.0 A 0.85 % + 1.8 A	

Parameter/Range	Frequency	CMC <sup>2,5</sup> (±)	Comments
AC Current <sup>3</sup> – Measure			
(10 to 100) µA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 10) kHz	0.40 % + 44 nA 0.15 % + 44 nA 0.061 % + 44 nA 0.061 % + 44 nA 0.061 % + 44 nA	Agilent 3458A
100 µA to 1 mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 10) kHz	0.40 % + 0.23 µA 0.15 % + 0.23 µA 0.061 % + 0.23 µA 0.031 % + 0.23 µA 0.061 % + 0.23 µA	
(1 to 10) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 10) kHz	0.40 % + 2.8 µA 0.15 % + 2.8 µA 0.061 % + 2.8 µA 0.031 % + 2.8 µA 0.061 % + 2.8 µA	
(10 to 100) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 10) kHz	0.40 % + 28 µA 0.15 % + 28 µA 0.061 % + 28 µA 0.031 % + 28 µA 0.061 % + 28 µA	
100 mA to 1 A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 10) kHz	0.40 % + 0.29 mA 0.18 % + 0.29 mA 0.081 % + 0.29 mA 0.11 % + 0.29 mA 0.31 % + 0.29 mA	
(1 to 3) A	10 Hz to 1 kHz (1 to 5) kHz	0.14 % + 3.8 mA 2.3 % + 4.5 mA	Fluke 8846A
(3 to 10) A	10 Hz to 1 kHz (1 to 5) kHz	0.14 % + 9.4 mA 2.3 % + 9.7 mA	
(10 to 20) A	10 Hz to 400 Hz 400 Hz to 1 kHz	0.23 % + 18 mA 0.23 % + 38 mA	Agilent 3458A & 50A/50mV current shunt

Parameter/Range	Frequency	CMC <sup>2, 5</sup> (±)	Comments
Capacitance <sup>3</sup> – Generate			
(0.19 to 0.4) nF	10 Hz to 10 kHz	0.39 % + 0.0079 nF	
(0.4 to 1.1) nF	10 Hz to 10 kHz	0.39 % + 0.0080 nF	
(1.1 to 3.3) nF	10 Hz to 3 kHz	0.39 % + 0.0085 nF	
(3.3 to 11) nF	10 Hz to 1 kHz	0.19 % + 0.012 nF	
(11 to 33) nF	10 Hz to 1 kHz	0.19 % + 0.080 nF	
(33 to 110) nF	10 Hz to 1 kHz	0.19 % + 0.10 nF	
(110 to 330) nF	10 Hz to 1 kHz	0.19 % + 0.32 nF	
(0.33 to 1.1) µF	(10 to 600) Hz	0.19 % + 1.1 nF	
(1.1 to 3.3) µF	(10 to 300) Hz	0.19 % + 3.0 nF	
(3.3 to 11) µF	(10 to 150) Hz	0.19 % + 12 nF	
(11 to 33) µF	(10 to 120) Hz	0.31 % + 23 nF	
(33 to 110) µF	(10 to 80) Hz	0.35 % + 150 nF	
(110 to 330) µF	Up to 50 Hz	0.35 % + 420 nF	
(0.33 to 1.1) mF	Up to 20 Hz	0.35 % + 1.1 µF	
(1.1 to 3.3) mF	Up to 6 Hz	0.35 % + 2.5 µF	
(3.3 to 11) mF	Up to 2 Hz	0.35 % + 9.5 µF	
(11 to 33) mF	Up to 0.6 Hz	0.58 % + 25 µF	
(33 to 110) mF	Up to 0.2 Hz	0.85 % + 81 µF	
(0.95 to 9.5) µF	1 kHz	0.70 % + 1.0 nF	Transmille 4010
(9.5 to 95) µF		0.70 % + 6.0 nF	
95 µF to 0.95mF		0.70 % + 5.3 nF	
(0.95 to 9.5) mF		0.70 % + 6.0 µF	
(9.5 to 100) mF		0.70 % + 24 µF	
Fixed Points Capacitance			
1 nF	1 kHz	0.25 % + 0.61 pF	
2 nF		0.25 % + 42 pF	
5 nF		0.25 % + 32 pF	
10 nF		0.25 % + 6.0 pF	
100nF		0.25 % + 60 pF	
1 µF		0.40 % + 0.60 µF	
10 µF		0.60 % + 0.60 µF	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of Thermocouple Indicating Systems <sup>3</sup> –			
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.34 °C 0.27 °C 0.24 °C 0.26 °C	Fluke 5520A
Type C	(0 to 150) °C (150 to 650) °C (650 to 1000) °C (1000 to 1800) °C (1800 to 2316) °C	0.23 °C 0.20 °C 0.24 °C 0.39 °C 0.65 °C	
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.39 °C 0.13 °C 0.11 °C 0.14 °C 0.17 °C	
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.21 °C 0.13 °C 0.11 °C 0.13 °C 0.18 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.26 °C 0.14 °C 0.13 °C 0.20 °C 0.31 °C	
Type L	(-200 to -100) °C (-100 to 800) °C (800 to 900) °C	0.29 °C 0.20 °C 0.14 °C	
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.31 °C 0.17 °C 0.15 °C 0.14 °C 0.21 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.44 °C 0.27 °C 0.26 °C 0.31 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of Thermocouple Indicating Systems <sup>3</sup> – (cont)			
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.36 °C 0.28 °C 0.29 °C 0.36 °C	Fluke 5520A
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.49 °C 0.19 °C 0.13 °C 0.11 °C	
Type U	(-200 to 0) °C (0 to 600) °C	0.43 °C 0.21 °C	
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.44 °C 0.34 °C 0.30 °C 0.33 °C	Transmille 4010
Type C	(0 to 150) °C (150 to 650) °C (650 to 1000) °C (1000 to 1800) °C (1800 to 2316) °C	0.30 °C 0.26 °C 0.31 °C 0.50 °C 0.84 °C	
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.50 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.27 °C 0.16 °C 0.17 °C 0.17 °C 0.23 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1370) °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.40 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of Thermocouple Indicating Systems <sup>3</sup> – (cont)			
Type L	(-200 to -100) °C (-100 to 800) °C (800 to 900) °C	0.37 °C 0.26 °C 0.17 °C	Transmille 4010
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.40 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.57 °C 0.35 °C 0.33 °C 0.40 °C	
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.47 °C 0.36 °C 0.37 °C 0.46 °C	
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.63 °C 0.24 °C 0.16 °C 0.14 °C	
Type U	(-200 to 0) °C (0 to 600) °C	0.56 °C 0.27 °C	
Electrical Simulation of RTD Indicators & Indicating Systems <sup>3</sup> –			
Pt 385, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C (400 to 630) °C	0.040 °C 0.040 °C 0.050 °C 0.070 °C 0.080 °C 0.090 °C 0.18 °C 0.090 °C	Fluke 5520A

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Simulation of RTD Indicators & Indicating Systems <sup>3</sup> – (cont)			
Pt 3926, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C	0.040 °C 0.040 °C 0.060 °C 0.070 °C 0.080 °C	Fluke 5520A
Pt 3916, 100 Ω	(-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.19 °C 0.030 °C 0.040 °C 0.050 °C 0.050 °C 0.060 °C 0.070 °C 0.080 °C 0.18 °C	
Pt 385, 200 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.030 °C 0.030 °C 0.030 °C 0.040 °C 0.090 °C 0.10 °C 0.11 °C 0.12 °C	
Pt 385, 500 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.030 °C 0.040 °C 0.040 °C 0.050 °C 0.060 °C 0.060 °C 0.070 °C 0.090 °C	
Pt 385, 1000 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.020 °C 0.020 °C 0.030 °C 0.040 °C 0.050 °C 0.050 °C 0.050 °C 0.18 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Electrical Simulation of RTD Indicators & Indicating Systems <sup>3</sup> —(cont)			
PtNi 385, 120 $\Omega$	(-80 to 0) °C (0 to 100) °C (100 to 260) °C	0.060 °C 0.060 °C 0.11 °C	Fluke 5520A
Cu 427, 10 $\Omega$	(-100 to 260) °C	0.23 °C	
Pt 25, 25 $\Omega$	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.50 °C 0.50 °C 0.60 °C 0.60 °C 0.60 °C 0.60 °C 0.60 °C	Transmille 4010
Pt 100, 100 $\Omega$	(-200 to -80) °C (80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.13 °C 0.13 °C 0.55 °C 0.55 °C 0.55 °C 0.55 °C 0.55 °C	
Pt 250, 250 $\Omega$	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 800) °C	0.25 °C 0.25 °C 0.30 °C 0.30 °C 0.30 °C 0.30 °C 0.30 °C	
Pt 25, 500 $\Omega$	(-200 to -80) °C (80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.10 °C 0.10 °C 0.90 °C 0.90 °C 0.90 °C 0.90 °C 0.90 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Electrical Simulation of RTD Indicators & Indicating Systems <sup>3</sup> – (cont)			
Pt 100, 1000 $\Omega$	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 800) °C	0.080 °C 0.080 °C 0.45 °C 0.45 °C 0.45 °C 0.45 °C 0.45 °C	Transmille 4010
Fixed Point Temperature (4 Wire)			
(-100 to 800) °C	-100 °C 0 °C 30 °C 60 °C 100 °C 200 °C 400 °C 800 °C	0.011 °C 0.002 °C 0.004 °C 0.008 °C 0.011 °C 0.022 °C 0.041 °C 0.08 °C	

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Oscilloscope <sup>3</sup> –  Square Wave Amplitude: 50 Ω Load 1 MΩ Load  DC Signal Output: 50 Ω Load 1 MΩ Load  Leveled Sine Wave Amplitude: 5 mV <sub>p-p</sub> to 5.5 V <sub>p-p</sub>  5 mV <sub>p-p</sub> to 3.5 V <sub>p-p</sub>  Flatness: Relative to 50 kHz 5 mV <sub>p-p</sub> to 5.5 V <sub>p-p</sub>  Time Marker: 50 Ω Load  Edge-Rise Time  Square Wave Amplitude: 1 MΩ Load  DC Signal Output: 1 MΩ Load	V <sub>p-p</sub> 1 mV to 6.6 V V <sub>p-p</sub> 1 mV to 130 V ≤ 1 kHz > 1 kHz  V <sub>dc</sub> (0 to 6.6) V V <sub>dc</sub> (0 to 130) V  50 kHz reference 50 KHz to 100 MHz (100 to 300) MHz  (300 to 600) MHz (600 to 1100) MHz  50 kHz reference 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (600 to 1100) MHz  5 s to 50 ms (20 to 1.0) ms (50 to 20) ns 10 ns (5 to 1) ns  1 kHz to 10 MHz  V <sub>p-p</sub> 1 mV to 300 V at 1 kHz  V <sub>dc</sub> (0 to 300) V	0.19 % of output + 39 µV 0.08 % of output + 39 µV 0.08 % of output + 39 µV  0.19 % of output + 5.9 µV 0.040 % of output + 8.0 µV  1.6 % of output + 250 µV 2.8 % of output + 400 µV 3.1 % of output + 660 µV  4.7 % of output + 660 µV 5.5 % of output + 250 µV  1.5 % of output + 240 µV 2.0 % of output + 240 µV 4.0 % of output + 240 µV 5.0 % of output + 240 µV  8.1 µs/s 8.1 µs/s 8.1 µs/s 5.8 µs/s 5.8 µs/s  ≤ 300 ps  0.25 % of output + 38 µV 0.01 % of output + 0.84 µV	Fluke 5520A with SC-1100 option  Transmille 4010 with SCP 600 option

Parameter/Equipment	Range	CMC <sup>2, 5</sup> (±)	Comments
Oscilloscope <sup>3</sup> – (cont)			
Leveled Sine Wave Amplitude: 600 mV <sub>p</sub>	50 kHz reference (> 5 to 600) MHz	0.26 % of output + 69 µV 0.26 % of output + 69 µV	Transmille 4010 with SCP 600 option
Timebase: 50 Ω Load	(5 to 1) s (500 to 1) ms (500 to 100) µs (50 to 1) µs (500 to 2) ns	0.010% + 0.58 ms 0.010 % + 58 µs 0.010% + 5.8 µs 0.010 % + 0.58 ns 0.010 % + 0.58 ps	
Edge-Rise Time	1 kHz	≤ 1.5 ns	

Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> (±)	Comments
Phase – Measuring Equipment <sup>3</sup>	(10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz  (40 to 60) Hz (60 to 100) Hz (100 to 400) Hz	0.093° 0.20° 0.39° 2.0° 3.9° 7.8°  0.11° 0.26° 0.50°	Fluke 5520A  Transmille 4010
DC Power <sup>3</sup> –			
33 mV to 1020 V	330 µA to 330 mA > 330 mA to 3 A (> 3 to 20) A	0.018 % 0.041 % 0.078 %	Fluke 5520A
200 mV to 1000 V	(0.2 to 20) mA > 20 mA to 2 A (> 2 to 30) A	0.030 % 0.014 % 0.031 %	Transmille 4010

Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
AC Power <sup>3</sup> –  PF = 1 33 mV to 1020 V			
(3 to 9) mA (> 9 to 33) mA (> 33 to 90) mA (> 90 to 330) mA > 330 mA to 0.9 A <td>(10 to 65) Hz</td> <td>0.062 % 0.041 % 0.063 % 0.042 % 0.047 % 0.066 % 0.11 % 0.10 %</br></td> <td>Fluke 5520A</td>	(10 to 65) Hz	0.062 % 0.041 % 0.063 % 0.042 % 0.047 % 	Fluke 5520A
(3 to 9) mA (> 9 to 33) mA (> 33 to 90) mA (> 90 to 330) mA > 330 mA to 0.9 A <td>(65 to 500) Hz</td> <td>0.049 % 0.046 % 0.049 % 0.074 % 0.055 % 0.069 % 0.11 % 0.12 %</br></td> <td></td>	(65 to 500) Hz	0.049 % 0.046 % 0.049 % 0.074 % 0.055 % 	
(> 3 to 9) mA (> 9 to 33) mA (> 33 to 90) mA (> 90 to 330) mA > 330 mA to 0.9 A <td>500 Hz to 1 kHz</td> <td>0.045 % 0.041 % 0.045 % 0.041 % 0.047 % 0.065 % 0.11 % 0.12 %</br></td> <td></td>	500 Hz to 1 kHz	0.045 % 0.041 % 0.045 % 0.041 % 0.047 % 	
(> 3 to 9) mA (> 9 to 33) mA (> 33 to 90) mA (> 90 to 330) mA > 330 mA to 0.9 A <td>(1 to 5) kHz</td> <td>0.10 % 0.068 % 0.10 % 0.084 % 0.47 % 0.47 %</br></td> <td></td>	(1 to 5) kHz	0.10 % 0.068 % 0.10 % 0.084 % 0.47 % 	
(> 3 to 9) mA (> 9 to 33) mA (> 33 to 90) mA (> 90 to 330) mA > 330 mA to 0.9 A <td>(5 to 10) kHz</td> <td>0.18 % 0.16 % 0.18 % 0.16 % 2.0 % 2.2 %</br></td> <td></td>	(5 to 10) kHz	0.18 % 0.16 % 0.18 % 0.16 % 2.0 % 	

Parameter/Range	Frequency	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
AC Power <sup>3</sup> – (cont)			
PF = 1 200 mV to 1000 V			
(0.2 to 20) mA > 20 mA to 2 A (> 2 to 30) A	(40 to 60) Hz	0.21 % 0.21 % 0.20 %	Transmille 4010
(0.2 to 20) mA > 20 mA to 2 A (> 2 to 30) A	(> 60 to 100) Hz	0.079 % 0.088 % 0.10 %	
(0.2 to 20) mA > 20 mA to 2 A (> 2 to 30) A	(> 100 to 400) Hz	0.079 % 0.088 % 0.31 %	

## II. Electrical – RF/Microwave

Parameter/Range	Frequency	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
RF Absolute Power – Measure			
(20 to 44) dBm	10 MHz to 4 GHz	0.73 dB	30 dB attenuator with HP8482B
(-30 to 20) dBm	0.1 MHz to 4 GHz	0.16 dB	HP437B & HP 8482A
(-20 to -70) dBm	10 MHz to 8 GHz (> 8 to 10) GHz (> 10 to 14) GHz (> 14 to 18) GHz	0.11 dB 0.13 dB 0.14 dB 0.16 dB	HP437B & HP8481D HP437B & HP8481D HP437B & HP8481D HP437B & HP8481D
(10 to -10) dBm	10 MHz to 8 GHz (> 8 to 10) GHz (> 10 to 14) GHz (> 14 to 18) GHz	0.13 dB 0.17 dB 0.17 dB 0.21 dB	HP437B & HP8481A HP437B & HP8481A HP437B & HP8481A HP437B & HP8481A

### III. Mechanical

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Pressure <sup>3</sup> – Generate and Measure			
Pneumatic	(-14 to 0) PSI	0.025 % + 0.73 PSI	Druck DPI610
	(0 to 300) PSI	0.025 % + 0.73 PSI	
Hydraulic	(0 to 2000) PSI	0.02 % + 0.34 PSI	Crystal Pressure 10KPSIXP2I and Additel ADT936
	(> 2000 to 10 000) PSI	0.1 % + 0.69 PSI	

### IV. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2, 7</sup> ( $\pm$ )	Comments
Frequency & Period – Measuring Equipment	10 MHz	0.035 mHz	Brandywine Communication GPS time/frequency system
	(0.01 to 119.99) Hz	3.0 pHz/Hz + 5.8 mHz	Fluke 5520A
	(120.0 to 1199.9) Hz	3.0 pHz/Hz + 58 mHz	
	(1.200 to 11.999) kHz	3.0 pHz/Hz + 0.58 Hz	(With 10 MHz GPS external clock)
	(12.00 to 119.99) kHz	3.0 pHz/Hz + 5.8 Hz	
	(120.0 to 1199.9) kHz	3.0 pHz/Hz + 58 Hz	
	(1.200 to 2.000) MHz	3.0 pHz/Hz + 0.58 kHz	
	(0.05 to 100) MHz	3.0 pHz/Hz + 5.8 kHz	
	(100 to 300) MHz	3.0 pHz/Hz + 5.8 kHz	
	(300 to 600) MHz	3.0 pHz/Hz + 5.8 kHz	
	(600 to 1100) MHz	3.0 pHz/Hz + 58 kHz	

## V. EMC

Parameter/Equipment	Range	CMC <sup>2, 4</sup> ( $\pm$ )	Comments
ESD Simulators –			EN 61000-4-2
Contact Voltage (Positive & Negative)	(2 to 8) kV $\pm$ 5 % (10 to 30) kV $\pm$ 5 %	0.35 % 4.4 %	Tektronix DPO7254 with Noiseken Target 06- 00067A, Noiseken Attenuator 00-00010A, Tektronix attenuator 011- 0053-03, Tektronix probe P6015A, Brandenburg HV meter 139D
Rise/Fall Time	(0.6 to 1) ns	1.2 %	
Peak Current	(7.5 to 30) A $\pm$ 15 %	2.3 %	
30 ns Current	(4 to 16) A $\pm$ 30 %	2.2 %	
60 ns Current	(2 to 8) A $\pm$ 30 %	3.0 %	
EFT/Burst Generator <sup>3</sup> – 50 $\Omega$ / 1 k $\Omega$ Load; 100 kHz / 5 kHz			EN 61000-4-4
Voltage (Positive & Negative)	(0.25 to 4.4) kV	4.3 %	Tektronix DPO7254 with Noiseken attenuators AT- 810 & AT-811
Rise/Fall Time	5 ns $\pm$ 30 %	3.6 %	
Burst Duration	50 ns $\pm$ 30 %	3.6 %	
Impulse Duration 5 kHz	20 ms $\pm$ 20 %	3.7 %	
100 kHz	1 ms $\pm$ 20 %	3.7 %	
Burst Period (5/100) kHz	300 ms $\pm$ 20 %	3.7 %	
Repetition Cycle 5 kHz	200 $\mu$ s $\pm$ 20 %	3.7 %	
100 kHz	10 $\mu$ s $\pm$ 20 %	3.7 %	

Parameter/Equipment	Range	CMC <sup>2, 4</sup> ( $\pm$ )	Comments
Surge Generator <sup>3</sup> –			EN 61000-4-5
Open Circuit – Voltage (Positive & Negative)	(0.5 to 4.4) kV	2.6 %	Tektronix DPO7254 with Pintek DP-100, Pearson 110A, Tektronix 011-0060-02
Rise/Fall Time	1.2 $\mu$ s	2.7 %	
Duration	50 $\mu$ s	2.6 %	
Short Circuit – Current (Positive & Negative)	(41.67 to 2200) A	1.3 %	
Front Time	8 $\mu$ s	1.3 %	
Duration	20 $\mu$ s	1.3 %	
Ring Wave Generator <sup>3</sup> –			EN 61000-4-12
Open Circuit – Voltage (Positive & Negative)	(0.5 to 4) kV	5.9 %	Tektronix DPO7254 with Pintek DP-100, Pearson 110A, Tektronix 011-0060-02
Rise/Fall Time	1 $\mu$ s	6.0 %	
Duration	10 $\mu$ s	6.0 %	
Short Circuit – Current (Positive * Negative)	(200 & 500) A	5.8 %	
Rise/Fall Time	8 $\mu$ s	6.0 %	

Parameter/Equipment	Range	CMC <sup>2, 4</sup> ( $\pm$ )	Comments
PQT <sup>3</sup> –			EN 61000-4-11
Testing Voltage at 220 V With no Loading –			
Dip Voltage	(0 to 70) %	2.7 %	Tektronix DPO7254 with Pintek DP-100
Duration (1 to 100) cycles	16.6 ms to 1.666 s	2.6 %	
Phase Angle	(0 to 360) $^{\circ}$	2.7 %	
Rise/Fall Time	2 cycles, 30 $^{\circ}$	2.7 %	

<sup>1</sup> This laboratory offers commercial and field calibration services.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> Unless otherwise noted, in the statement of CMC, percent refers to percent of reading.

<sup>5</sup> The measurands stated are generated using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure the measurand in the ranges indicated. CMC are expressed as either a specific value that covers the full range or as a fraction/percent of the reading plus a fixed floor specification.

<sup>6</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.

<sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.